


Introduction to CCD Astronomy

Jon Rees

Observational Astronomy Workshop

Astronomy By Eye

- Unaided limiting magnitude ~6
- Telescopes brought step-change
- But no direct record of observations, still limited on faint objects, optical illusions



Drawing of Jupiter by James Keeler, 1890 (Credit: Lick Observatory Historical Collections)

Astronomy By Eye


- Unaided limiting magnitude ~6
- Telescopes brought step-change
- But still difficult to deal with faint objects, optical illusions



Drawing of 'canals' on Mars by Percival Lowell, 1905 (Credit: Lowell Observatory)

Photographic Plates

- Stable, wide-field observations
- Excellent for large area surveys, e.g. Palomar, Schmidt
- Beyond visual wavelengths
- By exposing for long time - faint objects




Top: Negative (left) and Positive (right) prints of a 135 min exposure of the Pleiades, Dec 1898

Bottom: Photographic plate showing 4 hr exposure of an edge-on galaxy, Nov 1899

(Credit: Lick Observatory Historical Collections)


Photomultiplier Tubes

- Photons hit cathode, eject electrons, secondary electrodes amplify the effect
- Converts incident photons to electrical signal
- Linear response - Accurate calibration of photometry
- But only single element




Top: Layout of a Photomultiplier Tube (Credit: R. O'Connell, U.Va)

Bottom: V, B-V CMD of Praesepe (Johnson 1952)



The First CCD Observation


- Created by Bell Labs in 1969.
- First used for Astronomy in 1976 by JPL/UoA



1976 Observation of Uranus from the UoA 61-inch Telescope (Janesick & Blouke, 1987)

CCD Operation

- Doped semiconductor, photons liberate electrons
- Grid of electrodes -> potential wells (pixels)
- Voltages cycled to move charge to read-out amplifiers
- Conversion from analogue voltage to digital counts - ADC
- Gain is set by electronics, e/ADU



Cross section of 3-phase CCD & charge transfer diagram (Dawiec 2011)

File Format (FITS)


- Data stored in 'FITS' files
- FITS files start with ascii headers - contain useful information
- Data are stored in arrays after headers
- Many tools exist to read FITS e.g. DS9, IRAF, python routines

FITS Header for Skyflat		
Keyword	Value	Comment
SIMPLE	T	NORMAL FITS IMAGE
BITPIX	16	DATA PRECISION
NAXIS	2	NUMBER OF IMAGE DIMENSIONS
NAXIS1	1056	NUMBER OF COLUMNS
NAXIS2	1024	NUMBER OF ROWS
CRVAL1U	2048	COLUMN ORIGIN
CRVAL2U	2048	ROW ORIGIN
CDELTA1U	-2	COLUMN CHANGE PER PIXEL
CDELTA2U	-2	ROW CHANGE PER PIXEL
OBSNUM	1041	OBSERVATION NUMBER
IDNUM	2	IMAGE ID
UGEOM	0	UCAM READOUT GEOMETRY
DGEOM	0	DESCRAMBLE GEOMETRY
AMPSROW	1	AMPLIFIERS PER ROW
AMPSCOL	1	AMPLIFIERS PER COLUMN
OBSTYPE	'OBJECT'	IMAGE TYPE
EXPTIME	3	Exp time (not counting shutter error)
BSCALE	1	DATA SCALE FACTOR
BZERO	32768	DATA ZERO POINT
COMMENT		Real Value = FITS*BSCALE+BZERO
PROGRAM	'NEWCAM'	New Lick Camera
VERSION	'nickel_direct'	Data acquisition version
TSEC	1592624447	CLOCK TICK - SECONDS
TUSEC	656880	CLOCK TICK - MICROSECONDS
DATE	'2020-06-20T03:40:47.65'	UT of CCD readout & descramble
DATASEC	'[1:1024,1:1024]'	NOAO-style data section / IRAF
COMMENT		End of cards hard-coded in fits_cards
COMMENT		Begin of cards from other times
CSYER2	0.01666669920087	systematic error along direction of WCS axis i
CSYER1	0.01666669920087	systematic error along direction of WCS axis i
CRDER2	5.139999848325E-05	random error along direction of WCS axis i
CRDER1	5.139999848325E-05	random error along direction of WCS axis i
CD2_2	-0.0001027239995892	CTM element i_j from FITS axis j to WCS axis i
CD2_1	3.946270226152E-06	CTM element i_j from FITS axis j to WCS axis i
CD1_2	-3.946270226152E-06	CTM element i_j from FITS axis j to WCS axis i
CD1_1	-0.0001027239995892	CTM element i_j from FITS axis j to WCS axis i

Detector Characteristics

Sensitivity (QE)


- Quantum Efficiency - ability of detector to detect photons
- QE is a function of wavelength
- Detectors can be targeted at different wavelength regimes



Quantum efficiency for Nickel CCD2 (Credit: UCO/Lick)

Sensitivity (QE)




- Quantum Efficiency - ability of detector to detect photons
- QE is a function of wavelength
- Detectors can be targeted at different wavelength regimes



Photon absorption length in silicon (Reicke 1994)


Plate Scale/Binning

- Plate scale - relation between detector pixels and physical size on sky
- Can 'bin' groups of pixels together
- Decreases resolution, but improves readout time and readout noise

Binning Options	Combined pixels on the CCD Chip
None	
2 x 2 (4 pixels = 1)	
3 x 3 (9 pixels = 1)	

Linearity

- If detector was perfect, double number of photons \rightarrow double counts
- Only true up to a certain count limit
- At high counts, detectors may become non-linear



Saturation

- When electrons reach limit of ADC, no more can be counted
- Bright objects can cause electrons to exceed full well depth pixels
- Electrons will start to fill neighbouring pixels causing bleed trails



Electron bleed trails from saturated stars (Credit:ESO)

Read Noise

- Conversion from analog to digital signal introduces noise
- Electronics also introduce spurious electrons throughout readout
- Can often decrease read noise by using slower read out modes



Thermal Noise/Dark Current

- Thermal energy can liberate electrons
- These are indistinguishable from electrons liberated by photons
- Solution - cool the detector. Generally use liquid nitrogen
- Dark current negligible at these temperatures

Calibration Files


Bias

- Zero second exposure
- But signal isn't zero?
- We apply a constant voltage to the detector
- Positive base signal - prevent negative values
- Overscan vs Bias frame



Flat-Field

- Uniform illumination source
- Dome flats (easy) vs twilight sky flats (better)
- Shows non-uniformity of detector, along with e.g. dust, filter imperfections



Flat-Field

- Uniform illumination source
- Dome flats (easy) vs twilight sky flats (better)
- Shows non-uniformity of detector, along with e.g. dust, filter imperfections



Flat-Field

- Uniform illumination source
- Dome flats (easy) vs twilight sky flats (better)
- Shows non-uniformity of detector, along with e.g. dust, filter imperfections



Fringing



- Interference due to photons reflecting within CCD
- Occurs longwards of $\sim 700\text{nm}$
- Largely due to atmospheric OH - cannot correct with flats
- But largely stable with time - can use library frames to correct



Example Z-band fringe frame for INT-WFC


Cosmic Rays

- Blue - Relatively few events
- Red - Thicker chip, many cosmic ray events




Spectroscopy

- Same ideas apply to spectroscopy
- Bias/Flat fields
- Also arc lamps - wavelength calibration



Arcs



Conclusions

- CCDs are great!
- CCDs are not perfect
- Beware of non-linearity/saturation
- Remember calibration files

Conclusions

- Calibration Files:
 - Bias (Bias Voltage)
 - Flat Field (Non-uniform response)
 - Arcs (Wavelength Calibration)
 - Fringe Frame, Standard Star